

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 004931  
12:42 09/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:23  
PRESSURIZING 2:27  
STERILIZING 3:32  
136.5 C 221kPa 3:32  
136.4 C 225kPa 4:02  
136.3 C 223kPa 4:32  
136.3 C 223kPa 5:02  
136.2 C 223kPa 5:32  
136.3 C 221kPa 6:02  
136.4 C 222kPa 6:32  
136.4 C 224kPa 7:02  
136.4 C 224kPa 7:02  
VENTING 7:02  
AIR DRYING 7:54  
CYCLE COMPLETE 9:21

-----  
Digital Signature #  
0FE3A96E1DB5CC7C  
-----