
Midmark M11 - v1.0.5
Total Cycles: 3544

Name: RITTER2
Sterilizer ID: V2313228
04/22/2024 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.1 F 0.0
2:00 72.1 F 0.1
4:00 73.3 F 0.8
6:00 96.3 F 2.7
8:00 127.9 F 5.5
10:00 165.4 F 8.6
12:00 187.6 F 12.3
14:00 210.1 F 3.9
16:00 217.2 F 7.7
18:00 227.0 F 11.7
20:00 229.5 F 7.4
22:00 241.2 F 12.6
24:00 239.6 F 10.6
26:00 252.0 F 16.7
28:00 262.7 F 23.0
30:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.0 F 29.1
2:00 272.3 F 29.5
3:00 272.1 F 29.6
4:00 271.6 F 28.9

Min 271.4 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 230.9 F 0.0
10:00 202.3 F 0.0
15:00 197.7 F 0.0
20:00 224.7 F 0.0
25:00 216.6 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:19
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:08:26
CYCLE COMPLETE

04/22/2024 10:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____