
Midmark M11 - v1.0.5
Total Cycles: 3816

Name: RITTER2
Sterilizer ID: V2313228
07/26/2024 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.9 F 0.0
2:00 154.2 F 1.4
4:00 159.4 F 4.2
6:00 181.3 F 8.0
8:00 198.4 F 12.0
10:00 215.5 F 5.5
12:00 227.0 F 9.8
14:00 233.3 F 9.9
16:00 238.7 F 11.3
18:00 236.9 F 9.5
20:00 249.3 F 15.5
22:00 260.0 F 21.4
24:00 269.7 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.6
2:00 271.3 F 29.0
3:00 272.4 F 29.8
4:00 271.8 F 29.1

Min 271.3 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 235.7 F 0.0
10:00 223.6 F 0.0
15:00 213.0 F 0.0
20:00 204.5 F 0.0
25:00 202.1 F 0.0
30:00 203.2 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:02
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:03:46
CYCLE COMPLETE

07/26/2024 11:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____