
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001472
11:04 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:42
STERILIZING 4:13
136.4 C 224kPa 4:13
136.2 C 224kPa 4:43
136.4 C 226kPa 5:13
136.1 C 224kPa 5:43
136.0 C 224kPa 6:13
136.4 C 227kPa 6:43
136.4 C 226kPa 7:13
136.4 C 225kPa 7:43
136.4 C 225kPa 7:43
VENTING 7:44
AIR DRYING 8:37
CYCLE COMPLETE 11:04

Digital Signature #
1BB56C5BFCA54A36
