
Midmark M11 - v1.0.5
Total Cycles: 2908

Name: RITTER2
Sterilizer ID: V2313228
09/20/2023 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.2 F 0.0
2:00 161.3 F 1.7
4:00 175.4 F 5.1
6:00 194.6 F 9.2
8:00 211.7 F 10.8
10:00 221.4 F 7.7
12:00 235.5 F 12.5
14:00 234.5 F 9.7
16:00 236.9 F 9.4
18:00 246.4 F 13.9
20:00 257.9 F 20.0
22:00 268.0 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.2 F 29.7
2:00 272.0 F 29.4
3:00 272.9 F 30.0
4:00 271.8 F 29.3

Min 270.8 F 28.8
Max 273.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.9 F 0.0
5:00 205.7 F 0.0
10:00 193.2 F 0.0
15:00 190.8 F 0.0
20:00 186.0 F 0.0
25:00 183.6 F 0.0
30:00 181.7 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:13
CYCLE COMPLETE

09/20/2023 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____