

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003014  
9:59 25/01/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:46  
PRESSURIZING 3:47  
STERILIZING 4:50  
136.3 C 222kPa 4:50  
136.1 C 222kPa 5:20  
136.4 C 220kPa 5:50  
136.2 C 219kPa 6:20  
136.2 C 219kPa 6:50  
136.0 C 221kPa 7:20  
136.4 C 223kPa 7:50  
136.4 C 224kPa 8:20  
136.4 C 224kPa 8:20  
VENTING 8:20  
AIR DRYING 9:14  
CYCLE COMPLETE 9:15

-----  
Digital Signature #  
18441C34FBD01EF2  
-----