

-----  
Midmark M11 - v1.0.5  
Total Cycles: 635

Name: RITTER2  
Sterilizer ID: V2313228  
12/15/2021 10:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.6 F 0.0  
2:00 185.5 F 3.7  
4:00 198.4 F 9.1  
6:00 218.0 F 7.0  
8:00 227.7 F 10.0  
10:00 232.1 F 8.1  
12:00 241.5 F 11.6  
14:00 246.0 F 13.7  
16:00 258.9 F 20.8  
18:00 270.5 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.5 F 29.9  
2:00 272.1 F 29.2  
3:00 271.4 F 29.2  
4:00 272.5 F 29.9  
  
Min 271.3 F 28.9  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 215.9 F 0.0  
10:00 227.1 F 0.0  
15:00 224.3 F 0.0  
20:00 224.1 F 0.0  
25:00 226.9 F 0.0  
30:00 230.7 F 0.0  
  
Drying Complete

FILLING: 2:13  
HEATING: 18:51  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

12/15/2021 11:45 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_