
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001111
15:21 03/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:02
STERILIZING 3:05
136.0 C 220kPa 3:05
136.2 C 221kPa 3:35
136.3 C 225kPa 4:05
136.4 C 226kPa 4:35
136.4 C 225kPa 5:05
136.3 C 224kPa 5:35
136.4 C 225kPa 6:05
136.4 C 226kPa 6:35
136.4 C 226kPa 6:35
VENTING 6:35
AIR DRYING 7:24
CYCLE COMPLETE 7:34

Digital Signature #
BBDD39DD04162325
