
Midmark M11 - v1.0.0
Total Cycles: 3200

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2023 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.5 F 0.0
2:00 190.3 F 0.5
4:00 199.7 F 4.1
6:00 219.1 F 8.9
8:00 229.1 F 8.5
10:00 234.4 F 9.8
12:00 239.0 F 8.9
14:00 245.6 F 13.2
16:00 244.1 F 12.1
18:00 256.8 F 18.8
20:00 267.9 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.8
2:00 271.5 F 28.1
3:00 272.2 F 28.9
4:00 272.2 F 28.6

Min 271.3 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 217.0 F 0.0
10:00 237.5 F 0.0
15:00 236.5 F 0.0
20:00 223.6 F 0.0
25:00 224.6 F 0.0
30:00 210.2 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:20
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

05/04/2023 09:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____