
Midmark M11 - v1.0.5
Total Cycles: 1117

Name: RITTER2
Sterilizer ID: V2313228
05/25/2022 02:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.2 F 0.0
2:00 157.8 F 3.0
4:00 180.8 F 7.9
6:00 203.9 F 8.6
8:00 225.6 F 8.6
10:00 235.8 F 10.5
12:00 243.3 F 12.7
14:00 243.7 F 12.1
16:00 257.9 F 19.6
18:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 28.8
2:00 272.2 F 29.0
3:00 272.1 F 28.6
4:00 272.4 F 29.2

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.0 F 0.0
5:00 205.2 F 0.0
10:00 219.8 F 0.0
15:00 219.3 F 0.0
20:00 219.6 F 0.0
25:00 217.0 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 18:52
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

05/26/2022 06:50 AM
ACCEPTED: X
REJECTED:
Operator: U02