
Midmark M11 - v1.0.5
Total Cycles: 1176

Name: RITTER2
Sterilizer ID: V2313228
06/08/2022 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.6 F 0.0
2:00 169.4 F 3.2
4:00 184.0 F 7.2
6:00 203.9 F 12.3
8:00 217.4 F 7.2
10:00 233.2 F 12.6
12:00 235.5 F 10.0
14:00 235.4 F 8.7
16:00 248.7 F 15.3
18:00 260.5 F 22.0
20:00 271.4 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.5 F 29.9
2:00 271.5 F 29.2
3:00 271.9 F 29.9
4:00 272.4 F 29.8

Min 271.1 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 213.8 F 0.0
10:00 230.8 F 0.0
15:00 237.8 F 0.0
20:00 235.2 F 0.0
25:00 218.1 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

06/08/2022 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____