
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006008
10:28 19/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:37
STERILIZING 3:39
136.4 C 224kPa 3:39
136.3 C 224kPa 4:09
136.3 C 225kPa 4:39
136.3 C 225kPa 5:09
136.4 C 225kPa 5:39
136.2 C 224kPa 6:09
136.4 C 225kPa 6:39
136.4 C 223kPa 7:09
136.4 C 223kPa 7:09
VENTING 7:09
AIR DRYING 7:58
CYCLE COMPLETE 10:33

Digital Signature #
BB8B1876944E0588
