
Midmark M11 - v1.0.0
Total Cycles: 1647

Name: RITTER 1
Sterilizer ID: V2178198
02/17/2022 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.9 F 0.0
2:00 182.4 F 2.7
4:00 186.6 F 6.7
6:00 212.2 F 11.9
8:00 224.2 F 6.2
10:00 236.3 F 11.5
12:00 234.3 F 8.6
14:00 241.1 F 10.3
16:00 246.4 F 13.6
18:00 259.0 F 20.3
20:00 270.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.3
2:00 271.9 F 28.6
3:00 271.3 F 28.5
4:00 272.4 F 29.3

Min 271.2 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 237.9 F 0.0
10:00 234.5 F 0.0
15:00 222.5 F 0.0
20:00 220.0 F 0.0
25:00 227.2 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

02/17/2022 11:34 AM
ACCEPTED: X
REJECTED:
Operator: U02