

-----  
Midmark M11 - v1.0.0  
Total Cycles: 979

Name: RITTER 1  
Sterilizer ID: V2178198  
07/07/2021 04:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
2:00 185.5 F 3.0  
4:00 190.7 F 6.8  
6:00 211.3 F 11.5  
8:00 221.0 F 5.6  
10:00 228.6 F 10.3  
12:00 234.3 F 8.1  
14:00 243.6 F 12.8  
16:00 241.9 F 11.3  
18:00 253.9 F 17.4  
20:00 264.5 F 23.6  
22:00 271.3 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.6 F 29.0  
2:00 271.7 F 28.4  
3:00 271.5 F 28.6  
4:00 272.6 F 28.9

Min 271.1 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.6 F 0.0  
5:00 233.1 F 0.0  
10:00 227.7 F 0.0  
15:00 226.3 F 0.0  
Dry Cycle Aborted

FILLING: 1:36  
HEATING: 22:05  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 18:46  
TOTAL CYCLE: 00:49:59

07/07/2021 05:04 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01