
Midmark M11 - v1.0.0
Total Cycles: 1634

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2022 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 175.2 F 2.6
4:00 191.5 F 7.0
6:00 213.7 F 12.6
8:00 229.3 F 7.6
10:00 240.8 F 13.1
12:00 239.7 F 10.6
14:00 237.0 F 8.8
16:00 252.1 F 16.1
18:00 264.6 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 271.7 F 28.0
2:00 272.0 F 28.3
3:00 272.5 F 28.6
4:00 271.6 F 28.4

Min 271.3 F 27.9
Max 272.5 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 248.0 F 0.0
10:00 201.0 F 0.0
15:00 201.9 F 0.0
20:00 205.3 F 0.0
25:00 213.2 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

02/15/2022 11:31 AM
ACCEPTED: X
REJECTED:
Operator: U02