
Midmark M11 - v1.0.5
Total Cycles: 4442

Name: RITTER2
Sterilizer ID: V2313228
03/21/2025 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.0 F 0.0
2:00 146.4 F 1.7
4:00 150.3 F 4.5
6:00 175.9 F 8.1
8:00 188.9 F 12.2
10:00 216.5 F 6.1
12:00 228.1 F 10.2
14:00 232.2 F 8.8
16:00 239.1 F 11.6
18:00 237.0 F 9.7
20:00 248.8 F 15.6
22:00 259.2 F 21.5
24:00 269.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 272.7 F 30.3
1:00 272.1 F 30.2
2:00 271.8 F 30.3
3:00 271.7 F 30.4
4:00 271.8 F 30.3

Min 271.7 F 30.2
Max 272.7 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 226.5 F 0.0
10:00 239.3 F 0.0
15:00 224.1 F 0.0
20:00 210.3 F 0.0
25:00 202.1 F 0.0
30:00 198.6 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 25:47
STERILIZING: 4:00
VENTING: 3:59
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

03/21/2025 12:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____