
Midmark M11 - v1.0.0
Total Cycles: 5396

Name: RITTER 1
Sterilizer ID: V2178198
03/06/2025 06:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.8 F 0.0
2:00 112.0 F 1.0
4:00 156.5 F 3.1
6:00 180.5 F 6.4
8:00 180.8 F 10.1
10:00 219.4 F 7.1
12:00 212.6 F 5.2
14:00 225.0 F 8.9
16:00 235.2 F 13.1
18:00 235.8 F 9.4
20:00 240.3 F 10.4
22:00 244.2 F 12.6
24:00 255.3 F 18.3
26:00 265.5 F 24.5
28:00 270.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.7 F 29.7
2:00 271.5 F 29.1
3:00 272.5 F 29.8
4:00 271.9 F 29.2

Min 271.4 F 29.0
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 199.1 F 0.0
10:00 193.3 F 0.0
15:00 186.7 F 0.0
20:00 185.2 F 0.0
25:00 187.4 F 0.0
30:00 190.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 28:48
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:07:36
CYCLE COMPLETE

03/06/2025 07:19 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____