

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5193

Name: RITTER 1  
Sterilizer ID: V2178198  
12/16/2024 12:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.1 F 0.0  
2:00 167.2 F 0.1  
4:00 188.7 F 1.1  
6:00 205.7 F 5.5  
8:00 221.5 F 10.1  
10:00 228.1 F 7.3  
12:00 237.9 F 10.6  
14:00 234.5 F 7.6  
16:00 244.2 F 11.6  
18:00 246.8 F 13.2  
20:00 259.0 F 19.9  
22:00 269.8 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.0  
1:00 271.9 F 28.4  
2:00 272.7 F 28.8  
3:00 272.2 F 28.4  
4:00 271.7 F 27.9

Min 271.5 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.4 F 0.0  
5:00 239.3 F 0.0  
10:00 190.6 F 0.0  
15:00 186.8 F 0.0  
20:00 186.2 F 0.0  
25:00 185.7 F 0.0  
30:00 189.8 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 22:53  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:01:27  
CYCLE COMPLETE

12/16/2024 01:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_