

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1057

Name: RITTER2  
Sterilizer ID: V2313228  
05/02/2022 02:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

C232: FILL MODE  
WATER LOW  
CHAMBER TEMP 191F  
CHAMBER PRESS 0.0PSI  
\*\* ITEMS NOT STERILE \*\*

05/02/2022 02:06 PM