

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001883  
7:39 24/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:41  
STERILIZING 3:38  
136.0 C 220kPa 3:38  
136.0 C 220kPa 4:08  
136.2 C 223kPa 4:38  
136.1 C 219kPa 5:08  
136.4 C 223kPa 5:38  
136.0 C 221kPa 6:08  
136.4 C 224kPa 6:38  
136.5 C 225kPa 7:08  
136.5 C 225kPa 7:08  
VENTING 7:08  
AIR DRYING 7:57  
CYCLE COMPLETE 8:08

-----  
Digital Signature #  
D1096D15F1711AB7  
-----