

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001603  
15:24 16/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:49  
STERILIZING 3:58  
136.5 C 225kPa 3:58  
136.1 C 223kPa 4:28  
136.4 C 225kPa 4:58  
136.3 C 224kPa 5:28  
136.4 C 226kPa 5:58  
136.2 C 224kPa 6:28  
136.4 C 225kPa 6:58  
136.4 C 225kPa 7:28  
136.4 C 225kPa 7:28  
VENTING 7:29  
AIR DRYING 8:19  
CASSETTE REMOVED  
CYCLE COMPLETE 29:43  
-----

Digital Signature #  
39725FFDE6D27C9A  
-----