
Midmark M11 - v1.0.5
Total Cycles: 2155

Name: RITTER2
Sterilizer ID: V2313228
03/16/2023 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.6 F 0.0
2:00 152.6 F 2.7
4:00 164.8 F 5.9
6:00 188.7 F 10.1
8:00 211.5 F 6.0
10:00 222.3 F 8.5
12:00 235.3 F 13.4
14:00 236.3 F 10.4
16:00 235.0 F 8.4
18:00 248.0 F 14.9
20:00 259.3 F 21.1
22:00 269.5 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.8 F 29.8
2:00 271.6 F 29.1
3:00 272.4 F 30.0
4:00 272.2 F 29.5

Min 271.4 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 212.1 F 0.0
10:00 217.4 F 0.0
15:00 213.6 F 0.0
20:00 210.7 F 0.0
25:00 214.0 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:20
CYCLE COMPLETE

03/16/2023 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____