

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001605  
10:59 11/22/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:28  
PRESSURIZING 7:37  
STERILIZING 9:22  
136.0 C 220kPa 9:22  
136.5 C 221kPa 9:52  
136.1 C 218kPa 10:22  
136.4 C 221kPa 10:52  
136.4 C 220kPa 11:22  
135.9 C 220kPa 11:52  
136.4 C 221kPa 12:22  
136.4 C 222kPa 12:52  
VENTING 12:52  
AIR DRYING 14:00  
CYCLE COMPLETE 44:00

-----  
Digital Signature #  
1C419242E2469CF7  
-----