
Midmark M11 - v1.0.0
Total Cycles: 1708

Name: RITTER 1
Sterilizer ID: V2178198
03/10/2022 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 177.8 F 2.3
4:00 196.4 F 5.8
6:00 212.4 F 10.5
8:00 223.2 F 5.2
10:00 234.4 F 10.0
12:00 235.2 F 8.5
14:00 244.8 F 12.9
16:00 242.5 F 11.4
18:00 255.6 F 18.2
20:00 267.0 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.8 F 29.1
2:00 272.3 F 28.5
3:00 271.2 F 28.1
4:00 272.4 F 29.0

Min 271.1 F 28.0
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 218.5 F 0.0
10:00 219.2 F 0.0
15:00 230.6 F 0.0
20:00 225.4 F 0.0
25:00 237.2 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

03/10/2022 01:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____