
Midmark M11 - v1.0.0
Total Cycles: 1702

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2022 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.2 F 0.0
2:00 153.5 F 0.7
4:00 172.3 F 3.5
6:00 202.1 F 8.2
8:00 216.0 F 13.2
10:00 225.2 F 7.8
12:00 238.5 F 12.7
14:00 237.0 F 9.8
16:00 238.3 F 8.8
18:00 248.3 F 14.5
20:00 260.3 F 21.1
22:00 271.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.1
2:00 271.3 F 28.3
3:00 272.5 F 29.2
4:00 271.8 F 28.5

Min 271.3 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 234.7 F 0.0
10:00 238.0 F 0.0
15:00 228.3 F 0.0
20:00 228.3 F 0.0
25:00 229.0 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

03/09/2022 02:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____