
Midmark M11 - v1.0.0
Total Cycles: 4481

Name: RITTER 1
Sterilizer ID: V2178198
04/23/2024 12:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.0 F 0.0
2:00 86.2 F 0.4
4:00 115.8 F 2.1
6:00 138.0 F 4.5
8:00 176.7 F 7.3
10:00 187.4 F 10.4
12:00 209.6 F 8.9
14:00 213.4 F 4.9
16:00 223.6 F 8.3
18:00 232.8 F 12.2
20:00 233.2 F 8.1
22:00 243.5 F 13.0
24:00 241.0 F 10.8
26:00 252.1 F 16.3
28:00 262.2 F 22.1
30:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.2 F 29.0
2:00 272.8 F 29.4
3:00 271.5 F 29.0
4:00 272.3 F 29.1

Min 271.3 F 28.7
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.4 F 0.0
5:00 205.7 F 0.0
10:00 209.3 F 0.0
15:00 202.3 F 0.0
20:00 194.2 F 0.0
25:00 205.5 F 0.0
30:00 189.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:24
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:09:58
CYCLE COMPLETE

04/23/2024 02:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____