
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001405
15:32 25/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:40
STERILIZING 3:53
136.0 C 221kPa 3:53
136.1 C 221kPa 4:23
136.4 C 222kPa 4:53
135.8 C 220kPa 5:23
136.3 C 224kPa 5:53
136.5 C 225kPa 6:23
136.4 C 223kPa 6:53
136.4 C 224kPa 7:23
136.4 C 224kPa 7:23
VENTING 7:23
AIR DRYING 8:15
CYCLE COMPLETE 23:15

Digital Signature #
9E8C3CC5E435537D
