
Midmark M11 - v1.0.0
Total Cycles: 3956

Name: RITTER 1
Sterilizer ID: V2178198
10/30/2023 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.5 F 0.0
2:00 174.6 F 2.6
4:00 177.4 F 6.3
6:00 197.1 F 10.6
8:00 218.8 F 4.9
10:00 224.4 F 8.8
12:00 236.8 F 13.5
14:00 237.8 F 10.5
16:00 237.7 F 8.3
18:00 249.0 F 14.9
20:00 260.4 F 21.3
22:00 271.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.3
2:00 271.7 F 28.7
3:00 271.2 F 28.8
4:00 272.4 F 29.3

Min 271.0 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 222.6 F 0.0
10:00 234.0 F 0.0
15:00 229.8 F 0.0
20:00 229.6 F 0.0
25:00 223.0 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:47
CYCLE COMPLETE

10/30/2023 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____