
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002341
9:25 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:55
STERILIZING 3:57
136.0 C 220kPa 3:57
136.1 C 221kPa 4:27
136.0 C 221kPa 4:57
136.3 C 224kPa 5:27
136.3 C 224kPa 5:57
136.2 C 223kPa 6:27
136.3 C 224kPa 6:57
136.3 C 225kPa 7:27
136.3 C 225kPa 7:27
VENTING 7:27
AIR DRYING 8:21
CYCLE COMPLETE 14:21

Digital Signature #
F4F6B7CCFC6CC9EC
