
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001503
13:22 10/14/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:08
PRESSURIZING 6:41
STERILIZING 8:00
136.0 C 220kPa 8:00
136.5 C 223kPa 8:30
136.3 C 221kPa 9:00
136.1 C 221kPa 9:30
136.2 C 222kPa 10:00
135.8 C 220kPa 10:30
136.2 C 221kPa 11:00
136.4 C 223kPa 11:30
VENTING 11:30
AIR DRYING 12:35
CYCLE COMPLETE 42:35

Digital Signature #
32A0EB18A9B56E3A
