

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1658

Name: RITTER 1  
Sterilizer ID: V2178198  
02/23/2022 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.1 F 0.0  
2:00 153.7 F 1.5  
4:00 167.2 F 4.8  
6:00 190.2 F 9.2  
8:00 212.2 F 7.5  
10:00 223.5 F 7.6  
12:00 236.2 F 12.6  
14:00 236.5 F 9.6  
16:00 237.8 F 9.1  
18:00 248.1 F 14.4  
20:00 260.4 F 21.1  
22:00 271.5 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.6 F 28.9  
2:00 272.0 F 28.7  
3:00 272.5 F 28.8  
4:00 271.9 F 28.7

Min 271.6 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.2 F 0.0  
5:00 223.9 F 0.0  
10:00 232.9 F 0.0  
15:00 228.9 F 0.0  
20:00 223.0 F 0.0  
25:00 225.9 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:37  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:49  
CYCLE COMPLETE

02/23/2022 10:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_