
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002670
10:08 11/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:03
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:03

Digital Signature #
60DF5E71116CF933
