
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001253
10:33 12/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:17
PRESSURIZING 4:17
STERILIZING 5:56
136.4 C 224kPa 5:56
136.3 C 223kPa 6:26
136.4 C 224kPa 6:56
136.2 C 222kPa 7:26
136.4 C 225kPa 7:56
136.4 C 223kPa 8:26
136.2 C 223kPa 8:56
136.4 C 224kPa 9:26
136.4 C 224kPa 9:26
VENTING 9:26
AIR DRYING 10:23
CYCLE COMPLETE 25:23

Digital Signature #
20FE3FD4472E2D51
