
Midmark M11 - v1.0.5
Total Cycles: 3446

Name: RITTER2
Sterilizer ID: V2313228
03/08/2024 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.5 F 0.0
2:00 160.7 F 1.9
4:00 177.3 F 5.7
6:00 198.8 F 10.2
8:00 212.0 F 6.5
10:00 224.2 F 8.9
12:00 233.5 F 11.0
14:00 241.5 F 11.7
16:00 239.8 F 10.6
18:00 252.9 F 17.0
20:00 264.0 F 23.6
22:00 271.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.4 F 29.5
2:00 271.8 F 28.9
3:00 270.9 F 28.7
4:00 271.6 F 29.2

Min 270.9 F 28.5
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 231.8 F 0.0
10:00 224.1 F 0.0
15:00 222.3 F 0.0
20:00 221.8 F 0.0
25:00 223.5 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

03/08/2024 01:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____