
Midmark M11 - v1.0.0
Total Cycles: 3857

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2023 07:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 85.4 F 0.0
4:00 98.4 F 0.0
6:00 123.4 F 0.1
8:00 151.9 F 0.5
10:00 179.9 F 3.7
12:00 195.3 F 7.2
14:00 209.1 F 10.8
16:00 220.4 F 5.9
18:00 228.5 F 8.2
20:00 239.6 F 12.6
22:00 237.2 F 9.3
24:00 239.1 F 9.6
26:00 247.3 F 13.6
28:00 259.4 F 20.1
30:00 270.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.4 F 28.3
2:00 272.7 F 28.8
3:00 271.6 F 27.9
4:00 272.5 F 28.5

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 223.2 F 0.0
10:00 184.4 F 0.0
15:00 211.2 F 0.0
20:00 198.9 F 0.0
25:00 189.0 F 0.0
30:00 182.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:51
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:56
CYCLE COMPLETE

10/04/2023 08:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____