
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008337
8:52 24/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:34
STERILIZING 3:25
136.0 C 220kPa 3:25
136.4 C 222kPa 3:55
136.3 C 218kPa 4:25
136.4 C 222kPa 4:55
136.4 C 219kPa 5:25
136.3 C 222kPa 5:55
136.4 C 223kPa 6:25
136.4 C 223kPa 6:55
136.4 C 223kPa 6:55
VENTING 6:56
AIR DRYING 7:46
CYCLE COMPLETE 37:46

Digital Signature #
AA6837C4A03C3600
