
Midmark M11 - v1.0.5
Total Cycles: 839

Name: RITTER2
Sterilizer ID: V2313228
02/24/2022 02:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.1 F 0.0
2:00 176.8 F 3.1
4:00 192.5 F 7.8
6:00 227.3 F 13.2
8:00 230.0 F 8.5
10:00 237.3 F 12.0
12:00 239.8 F 11.7
14:00 240.5 F 10.8
16:00 253.8 F 17.5
18:00 265.2 F 24.4
20:00 272.0 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.5 F 29.4
2:00 271.4 F 29.3
3:00 272.0 F 29.2
4:00 272.0 F 29.4

Min 271.4 F 29.2
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 235.6 F 0.0
10:00 229.6 F 0.0
15:00 223.1 F 0.0
20:00 182.9 F 0.0
25:00 182.3 F 0.0
30:00 181.0 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

02/24/2022 03:51 PM
ACCEPTED: X
REJECTED:
Operator: U02