
Midmark M11 - v1.0.0
Total Cycles: 3046

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2023 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.2 F 0.0
2:00 172.5 F 3.5
4:00 179.7 F 7.6
6:00 200.8 F 12.4
8:00 218.5 F 6.5
10:00 231.6 F 11.6
12:00 232.6 F 8.2
14:00 241.3 F 11.1
16:00 245.4 F 12.9
18:00 258.0 F 19.7
20:00 269.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.1
2:00 271.5 F 28.3
3:00 272.2 F 29.2
4:00 272.2 F 28.8

Min 271.2 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 214.2 F 0.0
10:00 235.5 F 0.0
15:00 233.1 F 0.0
20:00 229.9 F 0.0
25:00 226.4 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

03/28/2023 02:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____