
Midmark M11 - v1.0.5
Total Cycles: 3109

Name: RITTER2
Sterilizer ID: V2313228
11/10/2023 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 75.2 F 0.0
4:00 91.8 F 0.3
6:00 105.1 F 1.6
8:00 139.8 F 4.4
10:00 168.0 F 7.5
12:00 190.0 F 10.9
14:00 209.9 F 3.7
16:00 215.7 F 7.3
18:00 227.8 F 11.2
20:00 228.3 F 6.8
22:00 240.8 F 12.5
24:00 239.4 F 10.6
26:00 251.9 F 16.8
28:00 262.9 F 23.2
30:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.6 F 29.7
2:00 271.1 F 28.9
3:00 272.1 F 29.8
4:00 271.6 F 29.0

Min 271.1 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 214.8 F 0.0
10:00 195.8 F 0.0
15:00 187.8 F 0.0
20:00 185.4 F 0.0
25:00 184.4 F 0.0
30:00 189.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:21
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:23
CYCLE COMPLETE

11/10/2023 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____