
Midmark M11 - v1.0.5
Total Cycles: 2689

Name: RITTER2
Sterilizer ID: V2313228
07/24/2023 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.1 F 0.0
2:00 174.0 F 1.0
4:00 181.3 F 4.6
6:00 203.7 F 8.6
8:00 224.2 F 12.9
10:00 226.8 F 7.6
12:00 235.5 F 12.4
14:00 235.8 F 9.8
16:00 240.0 F 9.7
18:00 247.9 F 14.6
20:00 259.3 F 20.9
22:00 269.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.4 F 29.6
2:00 272.1 F 29.0
3:00 271.7 F 29.3
4:00 272.4 F 29.4

Min 271.2 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 212.3 F 0.0
10:00 219.9 F 0.0
15:00 222.9 F 0.0
20:00 210.2 F 0.0
25:00 215.5 F 0.0
30:00 213.4 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 22:55
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:02:04
CYCLE COMPLETE

07/24/2023 03:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____