
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000585
9:40 06/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:13
STERILIZING 4:19
136.4 C 223kPa 4:19
136.4 C 225kPa 4:49
136.2 C 223kPa 5:19
136.4 C 224kPa 5:49
136.4 C 224kPa 6:19
136.4 C 222kPa 6:49
136.0 C 223kPa 7:19
136.3 C 223kPa 7:49
136.3 C 223kPa 7:49
VENTING 7:49
AIR DRYING 8:44
CYCLE COMPLETE 8:47

Digital Signature #
91C5A7DE7AADE58E
