

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3423

Name: RITTER2  
Sterilizer ID: V2313228  
02/29/2024 02:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.2 F 0.0  
2:00 135.1 F 1.1  
4:00 151.9 F 3.9  
6:00 177.6 F 8.0  
8:00 199.8 F 12.8  
10:00 220.3 F 6.3  
12:00 232.9 F 11.1  
14:00 231.7 F 7.7  
16:00 243.9 F 12.6  
18:00 244.0 F 12.5  
20:00 256.5 F 19.0  
22:00 267.6 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.4 F 29.1  
2:00 271.5 F 29.0  
3:00 272.0 F 29.0  
4:00 271.5 F 29.0

Min 271.2 F 28.6  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
5:00 236.6 F 0.0  
10:00 227.7 F 0.0  
15:00 226.5 F 0.0  
20:00 223.2 F 0.0  
25:00 223.7 F 0.0  
30:00 221.5 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 23:21  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:01:38  
CYCLE COMPLETE

02/29/2024 03:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_