
Midmark M11 - v1.0.0
Total Cycles: 1685

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2022 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.6 F 0.0
2:00 168.5 F 2.5
4:00 182.4 F 6.2
6:00 199.9 F 11.1
8:00 217.4 F 5.4
10:00 231.3 F 10.8
12:00 231.3 F 7.6
14:00 243.0 F 12.2
16:00 244.3 F 12.6
18:00 257.5 F 19.5
20:00 268.9 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 29.3
2:00 272.1 F 28.7
3:00 271.8 F 28.7
4:00 272.7 F 29.3

Min 271.5 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 222.4 F 0.0
10:00 225.5 F 0.0
15:00 213.7 F 0.0
20:00 213.3 F 0.0
25:00 220.6 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:08
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

03/01/2022 12:22 PM
ACCEPTED: X
REJECTED:
Operator: U02