
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002459
16:11 27/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:11
STERILIZING 4:15
136.0 C 222kPa 4:15
136.4 C 226kPa 4:45
136.2 C 222kPa 5:15
136.4 C 225kPa 5:45
136.4 C 226kPa 6:15
136.4 C 226kPa 6:45
136.4 C 226kPa 7:15
136.3 C 224kPa 7:45
136.3 C 224kPa 7:45
VENTING 7:45
AIR DRYING 8:41
CYCLE COMPLETE 8:45

Digital Signature #
CA0BDB9D1C89C411
