
Midmark M11 - v1.0.5
Total Cycles: 2239

Name: RITTER2
Sterilizer ID: V2313228
04/12/2023 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 187.6 F 3.8
4:00 197.9 F 9.0
6:00 219.6 F 8.0
8:00 231.2 F 10.5
10:00 233.1 F 8.3
12:00 239.4 F 10.4
14:00 247.7 F 14.7
16:00 260.5 F 21.8
18:00 271.9 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 272.1 F 29.7
2:00 272.3 F 29.4
3:00 271.4 F 29.1
4:00 272.7 F 29.9

Min 271.3 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 225.0 F 0.0
10:00 227.8 F 0.0
15:00 227.6 F 0.0
20:00 227.1 F 0.0
25:00 227.8 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 18:26
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:34
CYCLE COMPLETE

04/12/2023 11:51 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____