
Midmark M11 - v1.0.0
Total Cycles: 3610

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2023 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.7 F 0.0
2:00 187.2 F 0.6
4:00 197.3 F 4.7
6:00 223.0 F 10.5
8:00 224.7 F 5.5
10:00 237.2 F 11.2
12:00 235.2 F 8.5
14:00 240.5 F 10.0
16:00 247.9 F 14.1
18:00 260.5 F 21.1
20:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 271.9 F 28.3
2:00 272.5 F 28.8
3:00 272.2 F 29.0
4:00 271.8 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 216.2 F 0.0
10:00 231.0 F 0.0
15:00 226.8 F 0.0
20:00 230.1 F 0.0
25:00 225.1 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

08/03/2023 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____