

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3866

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2023 06:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
2:00 184.9 F 3.5  
4:00 203.8 F 8.5  
6:00 222.4 F 10.5  
8:00 229.8 F 8.1  
10:00 241.3 F 13.1  
12:00 239.6 F 10.6  
14:00 236.8 F 8.9  
16:00 251.4 F 15.8  
18:00 263.0 F 22.5  
20:00 272.2 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.1 F 28.4  
2:00 272.2 F 28.8  
3:00 271.6 F 28.1  
4:00 272.4 F 28.8  
  
Min 271.1 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 237.8 F 0.0  
10:00 230.3 F 0.0  
15:00 222.5 F 0.0  
20:00 197.9 F 0.0  
25:00 190.8 F 0.0  
30:00 186.2 F 0.0  
Drying Complete

FILLING: 2:13  
HEATING: 20:09  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:00:02  
CYCLE COMPLETE

10/04/2023 07:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_