
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000812
16:20 23/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:19
PRESSURIZING 7:31
STERILIZING 9:22
136.4 C 222kPa 9:22
136.3 C 222kPa 9:52
136.2 C 221kPa 10:22
136.4 C 223kPa 10:52
136.1 C 220kPa 11:22
136.4 C 223kPa 11:52
136.4 C 223kPa 12:22
136.4 C 222kPa 12:52
136.4 C 222kPa 12:52
VENTING 12:52
CYCLE ABORTED 13:18
STOP BUTTON PRESSED
NOT STERILE
END TIME 13:18

Digital Signature #
22CC6CD1908031FB
