
Midmark M11 - v1.0.5
Total Cycles: 931

Name: RITTER2
Sterilizer ID: V2313228
03/30/2022 02:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.3 F 0.0
2:00 194.4 F 2.1
4:00 203.3 F 7.7
6:00 223.4 F 13.4
8:00 232.8 F 10.3
10:00 233.1 F 8.3
12:00 239.1 F 10.2
14:00 248.9 F 15.2
16:00 261.8 F 22.5
18:00 272.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.3 F 29.7
2:00 272.0 F 29.1
3:00 271.1 F 28.6
4:00 271.5 F 29.4

Min 271.0 F 28.6
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 206.4 F 0.0
10:00 229.9 F 0.0
15:00 234.9 F 0.0
20:00 228.8 F 0.0
25:00 223.5 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 18:20
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:11
CYCLE COMPLETE

03/30/2022 03:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____