
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 000304
15:46 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:37
STERILIZING 3:59
136.4 C 233kPa 3:59
136.2 C 232kPa 4:29
136.0 C 231kPa 4:59
136.4 C 233kPa 5:29
136.2 C 232kPa 5:59
136.4 C 233kPa 6:29
136.2 C 231kPa 6:59
136.4 C 233kPa 7:29
136.4 C 233kPa 7:29
VENTING 7:29
AIR DRYING 8:24
CYCLE COMPLETE 23:24

Digital Signature #
BC3D97EDA0B16D6C
